

Australian Standard<sup>®</sup>

---

**Screwcap bottles for  
microbiological use  
(McCartney bottles)**

---

This Australian Standard was prepared by Committee CH/1, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 1 March 1994 and published on 11 July 1994.

---

The following interests are represented on Committee CH/1:

Australian Chamber of Commerce and Industry  
Australian Government Analytical Laboratories  
National Standards Commission  
N.S.W. Agriculture  
National Association of Testing Authorities, Australia  
Royal Australian Chemical Institute  
Royal College of Pathologists of Australasia  
Royal Melbourne Institute of Technology  
University of Sydney

---

**Review of Australian Standards.** To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

---

*This Standard was issued in draft form for comment as DR 83209.*

Australian Standard<sup>®</sup>

---

**Screwcap bottles for  
microbiological use  
(McCartney bottles)**

---

First published as AS 4186—1994.

## PREFACE

This Standard was prepared by the Standards Australia Committee on Laboratory Glassware and Related Apparatus in response to a request from N.S.W. Agriculture supported by the Australian Society of Microbiologists and the N.S.W. Supply Service. The N.S.W. Supply Service has withdrawn a thin-walled variation of the McCartney bottle as a result of complaints received about its narrow neck, instability, leakage problems, tendency to crack during flame-sterilization procedures and soft cap which distorted under light loads.

In preparing this Standard, the Committee considered ISO 4796, *Laboratory glassware—Bottles*, and details given in Cruickshank, R. *Medical Microbiology—A guide to the laboratory diagnosis and control of infection*, 11th ed. Edinburgh: Livingstone, 1965, p. 724–734.

The terms ‘normative’ and ‘informative’ have been used in this Standard to define the application of the appendix to which they apply. A ‘normative’ appendix is an integral part of a Standard, whereas an ‘informative’ appendix is only for information and guidance.

## CONTENTS

	<i>Page</i>
1 SCOPE .....	3
2 REFERENCED DOCUMENTS .....	3
3 DEFINITIONS .....	3
4 MATERIALS .....	3
5 DESIGN REQUIREMENTS .....	4
6 TEST REQUIREMENTS .....	5
7 MARKING .....	5

## APPENDICES

A MEANS FOR DEMONSTRATING COMPLIANCE WITH THIS STANDARD .....	6
B DETERMINATION OF RESISTANCE TO THERMAL SHOCK .....	8
C EFFECT OF INSERT ON MICROBIAL GROWTH .....	10
D DETERMINATION OF RESISTANCE TO CORROSION .....	12
E DETECTION OF STRAIN IN GLASS .....	13
F DETERMINATION OF RESISTANCE TO LEAKAGE .....	15

### © Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

## STANDARDS AUSTRALIA

**Australian Standard****Screwcap bottles for microbiological use (McCartney bottles)**

**1 SCOPE** This Standard specifies the requirements for screwcap glass bottles suitable for use in microbiological laboratories as culture bottles.

**NOTES:**

- 1 Clause 6 specifies type tests that are performed to assess compliance of finished product with this Standard. However, it may also be convenient for product manufacturers to use these tests for quality control purposes.
- 2 Alternative methods for determining compliance of a lot with this Standard are given in Appendix A.

**2 REFERENCED DOCUMENTS** The following documents are referred to in this Standard:

**AS**

- |        |  |
|--------|--|
| 1199   | Sampling procedures and tables for inspection by attributes                  |
| 1399   | Guide to AS 1199—Sampling procedures and tables for inspection by attributes |
| 2243   | Safety in laboratories   |
| 2243.3 | Part 3: Microbiology   |
| 3900   | Quality management and quality assurance standards                           |
| 3900.1 | Part 1: Guidelines for selection and use                                     |
| 3904   | Quality management and quality system elements                               |
| 3904.1 | Part 1: Guidelines   |

**SAA**

- |          |   |
|----------|---|
| HB 18    | Guidelines for third-party certification and accreditation  |
| HB 18.44 | Guide 44: General rules for ISO or IEC international third-party certification schemes for products |

**3 DEFINITIONS** For the purpose of this Standard, the definitions below apply.

**3.1 Nominal capacity**—volume of liquid which a bottle will contain when filled to the base of the neck.

**3.2 Seal**—the system which enables the contents of the bottle to be isolated from the external environment.

**3.3 Transition zone**—curved part of the bottle where a change in direction occurs in the linear outline of the bottle, for example between base and sidewall or sidewall and neck.

**4 MATERIALS**

**4.1 Bottle** The bottle shall be made from clear, colourless glass free of particulate matter and having thermal properties which meet the requirements of the test outlined in Appendix B.

**4.2 Cap** The cap shall be made of material which is resistant to the conditions of normal use, including washing and sterilization procedures.

The cap shall be sufficiently strong and rigid so as not to crack or distort when being screwed onto the bottle.

**4.3 Insert** The insert shall be made of an elastic material which is chemically inert and retains its shape under normal conditions of use.

NOTE: Sealing inserts may be moulded into the cap.