AS 3778.3.4—1990 ISO 1088: 1985

Australian Standard®

Measurement of water flow in open channels

Part 3: Velocity-area methods Method 3.4: Collection and processing of data for determination of errors in measurement This Australian Standard was prepared by Committee CE/24, Measurement of Water Flow in Open Channels and Closed Conduits. It was approved on behalf of the Council of Standards Australia on 9 April 1990 and published on 10 December 1990.

The following interests are represented on Committee CE/24:

Association of Consulting Engineers of Australia

Australian Water and Wastewater Association

Board of Works, Melbourne

Department of Water Resources, NSW

Engineering and Water Supply Department of South Australia

Forestry Commission, NSW

Institute of Instrumentation and Control

Monash University

Public Works Department, NSW

Snowy Mountains Engineering Corporation

University of New South Wales

University of Queensland

Water Authority of Western Australia

Water Board, Sydney

Water Resources Commission, Queensland

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto. Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard®

Measurement of water flow in open channels

Part 3: Velocity-area methods Method 3.4: Collection and processing of data for determination of errors in measurement

First published as AS 3778.3.4-1990.

Incorporating Amdt 1—1991

PUBLISHED BY STANDARDS AUSTRALIA (STANDARDS ASSOCIATION OF AUSTRALIA) 1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 6298 6

PREFACE

This Standard was prepared by the Standards Australia Committee on Measurement of Water Flow in Open Channels and Closed Conduits. It is identical with and has been reproduced from ISO 1088—1985, *Liquidflow measurement in open channels*—*Velocity-areamethods*—*Collection and processing of data for determination of errors in measurement.*

This Standard is one of a series which deals with methods of measurement of water flow in open channels. The series when complete will consist of the following parts:

- Part 1: Vocabulary and symbols
- Part 2.1: General—Guidelines for the selection of methods of measurement
- Part 2.2: General—Establishment and operation of a gauging station
- Part 2.3: General—Determination of the stage-discharge relation (this Standard)
- Part 2.4: General—Estimation of uncertainty of a flow-rate measurement
- Part 2.5: General—Guidelines for the selection of flow gauging structures
- Part 3: Velocity-area methods
- Method 3.1: Measurement by current-meters and floats
- Method 3.2: Measurement by moving-boat method
- Method 3.3: Measurement by slope-area method
- Method 3.4: Collection and processing of data for determination of errors in measurement (this Standard)
- Method 3.5: Investigation of total error
- Method 3.6: Measurement of flow in tidal channels
- Method 3.7: Measurement by ultrasonic (acoustic) method
- Method 3.8: Electromagnetic method using a full-channel-width coil
- Part 4: Measurement structure methods
- Method 4.1: Thin-plate weirs
- Method 4.2: Rectangular broad-crested weirs
- Method 4.3: Round-nose horizontal broad-crested weirs
- Method 4.4: V-shaped broad-crested weirs
- Method 4.5: Triangular profile weirs
- Method 4.6: Flat-V weirs
- Method 4.7: Rectangular, trapezoidal and U-shaped flumes
- Method 4.8: Trapezoidal profile weirs
- Method 4.9: Parshall and Saniiri flumes
- Method 4.10: End-depth method for estimation of flow in rectangular channels with a free overfall
- Method 4.11: End-depth method for estimation of flow in non-rectangular channels with a free overfall (approximate method)
- Part 5: Dilution methods
- Method 5.1: Constant-rate injection method for the measurement of steady flow
- Method 5.2: Integration method for the measurement of steady flow
- Part 6.1: Measuring devices, instruments and equipment—Rotating element current-meters
- Part 6.2: Measuring devices, instruments and equipment—Direct depth sounding
- Part 6.3: Measuring devices, instruments and equipment—Calibration of rotating element current-meters in straight open tanks
- Part 6.4: Measuring devices, instruments and equipment—Echo sounders for water depth measurements
- Part 6.5: Measuring devices, instruments and equipment—Water level measuring devices
- Part 6.6: Measuring devices, instruments and equipment—Cableway system for stream gauging
- Part 6.7: Measuring devices, instruments and equipment—Ultrasonic (acoustic) velocity meters
- Part 6.8: Measuring devices, instruments and equipment—Position fixing equipment for hydrometric boats
- For the purposes of this Australian Standard, the ISO text should be modified as follows:
- (a) Wherever the words 'InternationalStandard' appear referring to this Standard, they should be read as 'Australian Standard'.
- (b) Wherever the word 'fluid' appears, it should be read as 'water'.
- (c) Substitute a point (.) for a comma (,) as a decimal marker.
- (d) The references to other publications should be replaced by references to Australian Standards as follows.

Reference to International Standard ISO

- 772 Liquid flow measurement in open channels—Vocabulary and symbols5168 Measurement of fluid flow—Estimation of
- uncertainty of a flow-rate measurement 748 Liquid flow measurement in open
- channels—Velocity-area methods

Australian Standard

AS

- 3778 Measurement of water flow in open channels
- 3778.1 Part 1: Vocabulary and symbols
- 3778.2.4 Part 2.4: General estimation of uncertainty of a flow-rate measurement
- 3778.3 Part 3: Velocity-area methods
- 3778.3.1 Method 3.1: Measurement by currentmeters and floats

© Copyright - STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

Part 3: Velocity-area methods Method 3.4: Collection and processing of data for determination of errors in measurement

1 Scope and field of application

This International Standard specifies a standard basis for the collection and processing of data for the determination of individual components of the total error in the measurement of liquid flow in open channels by velocity-area methods.

For determining the discharge in open channels by the velocity-area method, components of the flow need to be measured. The total uncertainty in discharge is a combination of the uncertainties in these components. This International Standard specifies a standard basis for collecting and processing the data required to compute the component uncertainties for determining the total uncertainty in discharge. This International Standard may be used when carrying out an investigation of component uncertainties from a large sample of rivers in a basin or in a country or for international investigations.

2 References

ISO 748, Liquid flow measurement in open channels – Velocity-area methods.

ISO 772, Liquid flow measurement in open channels – Vocabulary and symbols.

ISO 4363, Liquid flow measurement in open channels – Methods for measurement of suspended sediment.

ISO 4364, Liquid flow measurement in open channels – Bed material sampling.

ISO 5168, Measurement of fluid flow – Estimation of uncertainty of a flow-rate measurement.

ISO/TR 7178, Liquid flow measurement in open channels – Velocity-area methods – Investigation of total error.

3 General

3.1 Principle

The principle of the velocity-area method consists in determining from measurements the distribution of the flow velocity in the cross-section and the cross-sectional area, and using these observations for the computation of the discharge.

The measurements of the flow velocity are made in a number of verticals. In each vertical the mean velocity is determined from measurements at a selected number of points. The discharge per unit width may be found by multiplying the mean velocity by the depth in the vertical considered.

Each vertical is assumed to be representative of a segment of the cross-sectional area. The selection of the number and location of the verticals determines the width of these segments. Assuming that the discharge has remained constant during the measurements, summation of the discharge in the various segments gives the total discharge through the section.

3.2 Occurrence of error

When measuring width, depth and flow velocity, errors occur. The application of certain computational methods also introduces errors depending on the assumptions made.

A distinction shall be made between random and systematic errors, resulting from the instruments used, the measuring procedures and the processing of data. Random errors are also influenced by the nature of turbulent flow. The magnitude of random errors can be influenced favourably by the proper selection of instruments and methods. Systematic errors may be constant or variable and they cannot be eliminated by repeating the measurements or by increasing the duration of a measurement. There are, in addition, mistakes due to misreading an instrument or to instrument malfunction.